


## DAICEL CHIRAL TECHNOLOGIES GLOBAL APPLICATION CENTRE

### Avacopan

#### CHROMATOGRAPHIC CONDITIONS

Column: CHIRALPAK® IM [4.6 x 250mm, 5µm] Part No: 92325  
Mobile Phase: n-Hexane: DCM: Ethanol: DEA (80/05/15/0.1, v/v/v/v)  
Column Oven Temperature: 15°C  
UV Detection: 250 nm  
Flow Rate: 1.0 ml/min  
Diluent: n-Hexane: Ethanol (50/50)  
Sample Concentration: 1.0 mg/ml

Avacopan



#### SEPARATION PERFORMANCE

$tR_1$ : 5.3	$N_1$ : 2115	$T_1$ : 1.2	
$tR_2$ : 6.9	$N_2$ : 4361	$T_2$ : 1.2	$R_s$ : 3.5

